

REQUEST FOR QUOTATION LOCAL PURCHASE SECTION

RFQ NUMBER : 2492200354
DATE : 20.06.2022
REQ No : 1172200397
REQ TITLE : Supply and Instalation of Po
DELIVERY LOCATION
CLOSING DATE : 05.07.2022
CLOSING TIME : 10:00:00

| SL NO | ITEM CODE | SERVICE ID | DESCRIPTION | QTY | UOM | BRAND/ORIGIN | WARRANTY | UNIT PRICE | TOTAL PRICE |
|-------|------------|------------|--|-----|-----|--------------|----------|------------|-------------|
| 1 | 7000000049 | | MANUFACTURER = BIOLOGIC MODEL = SP-300 POTENTIOSTAT TECHNICAL SPECIFICATIONS= VOLTAGE # COMPLIANCE: ± 12 V ; ± 30 V WITH 2A/30V BOOSTE R # CONTROL VOLTAGE: ± 10 V # VOLTAGE RESOLUTION: 1 #V ON 60 MV RANGE CURRENT # CURRENT RANGES: 500 MA TO 10 NA (STANDARD); DOWN TO 1 PA (ULTRA LOW CURRENT) # MAXIMUM CURRENT: ± 500 MA (STANDARD); UP TO 2A WITH 2A/30V BOOSTER # CURRENT RESOLUTION: 760 FA (STANDARD) # LOW CURRENT: 6 RANGES FROM 100 NA TO 1 PA WITH RESOLUTION TO 76 AA EIS # FREQUENCY RANGE: 7 MHZ (3%, 3°) DOWN TO 10 #HZ; 3 MHZ (1 %, 1°) # EIS QUALITY INDICATORS THE FOLLOWING TO BE INCLUDED: [094-100] SP-300 CHASSIS - 2 SLOTS (2 CHANNELS MAX) - NO CHANNEL - IN CLUDING CALIBRATION BOARD & EC-LAB SOFTWARE PACKAGE (CHANNEL BOARD TO FIX AT SP-300 CHASSIS) [094-081/2] POTENTIOSTAT/GALVANOSTAT B OARD WITH EIS(/Z) OPTION FOR SP-200/SP-240/SP- | 1 | EA | | | | |

REQUEST FOR QUOTATION LOCAL PURCHASE SECTION

RFQ NUMBER : 2492200354
DATE : 20.06.2022
REQ No : 1172200397
REQ TITLE : Supply and Instalation of Po
DELIVERY LOCATION
CLOSING DATE : 05.07.2022
CLOSING TIME : 10:00:00

| SL NO | ITEM CODE | SERVICE ID | DESCRIPTION | QTY | UOM | BRAND/ORIGIN | WARRANTY | UNIT PRICE | TOTAL PRICE |
|-------|-----------|------------|--|-----|-----|--------------|----------|------------|-------------|
| | | | <p>300 (WITHOUT CELL CABLE) - USES ONE SLOT - +/-12 V +/-500 MA (BOOSTER BOARD TO FIX A T SP300 CHASIS) [094-081/10] 'SP-300 ±30V/2A BOOSTER [094-081/3] STANDARD CELL CABLE WITH ELECTROMETER FOR SP-200/240/300 POTENTIOST AT BOARD (1.75 METER) [094-081/6] LINEAR SCAN GENERATOR OPTION FOR SP-200/240/300/VSP-300/VMP-300 [092-1001/40] BAG OF 2 MM ALLIGATO R CLIPS (8X RED / BLUE / WHITE, 4X BLACK) [092-1001/44] BAG OF 2 MM BANANA PLUGS (10X RED / BLUE / BLACK / WHITE) [094-081/4] CABLE WITH ULTRA LOW CURRENT ELECTROMETER FOR SP-200/SP-240/SP-300 - USES NO SLOT (1.75 METER) [094-084/1] FC-45 FARADAY CAGE [INST-COMM]</p> <p>INSTALLATION & COMMISSIONING - ONE DAY BY BIOLOGIC MIDDLE EAST ENGINEER, TO BE INCLUDED DELIVERY TO DEWA R&D IN AL QUDRA, TO BE INC LUDED. INSTALLATION, COMISSIONING & TRAINING TO BE INCLUDED. "DEWA R&D LABORATORY CONTRACTORS/SUPPLIERS REQUIREMENTS</p> | | | | | | |

REQUEST FOR QUOTATION LOCAL PURCHASE SECTION

RFQ NUMBER : 2492200354
DATE : 20.06.2022
REQ No : 1172200397
REQ TITLE : Supply and Instalation of Po
DELIVERY LOCATION
CLOSING DATE : 05.07.2022
CLOSING TIME : 10:00:00

| SL NO | ITEM CODE | SERVICE ID | DESCRIPTION | QTY | UOM | BRAND/ORIGIN | WARRANTY | UNIT PRICE | TOTAL PRICE |
|-------|-----------|------------|---|-----|-----|--------------|----------|------------|-------------|
| | | | CHECKLIST" TO B E FILLED AND PROVIDED. SITE ASSESSMENT BEFORE DELIVERY IS MANDATORY BY THE SUPPLIER. DOCUMENTATIONS: THE USER MANUALS, MAINTENANCE P ROCEDURES, OPERATING PROCEDURES (NORMAL AND EMERGENCY OPERATING PROCEDURES IF APPLICABLE), TRAINING MATERIAL, DATASHEETS, CERTIFICAT ES OR ANY OTHER DOCUMENT RELATED TO THE EQUIPMENT SHOULD BE PROVIDED DURING THE COMMISSIONING IN HARD AND SOFT COPIES. | | | | | | |

TOTAL AMOUNT IN WORDS:

TOTAL AMOUNT:

SPECIAL NOTES / INSTRUCTIONS:

Supply and Instalation of Portable potentiostat/galvanostat biologic
instrument for electrochemical workstation for advanced materials lab

STANDARD TERMS & CONDITIONS

1) Prices should be 'DDP' delivery duty paid at DEWA stores.

MATERIAL AND OR SERVICE PROVIDED TO DEWA SHOULD PREFERABLY BE ENERGY EFFICIENT AND ENVIRONMENT FRIENDLY.

Page3 of 4

REQUEST FOR QUOTATION LOCAL PURCHASE SECTION

RFQ NUMBER : 2492200354
DATE : 20.06.2022
REQ No : 1172200397
REQ TITLE : Supply and Instalation of Po
DELIVERY LOCATION
CLOSING DATE : 05.07.2022
CLOSING TIME : 10:00:00

- 2) Quotation to be submitted only in local currency U.A.E Dirhams
- 3) DEWA Standard payment terms is '30 days credit' from the date of acceptance of material
- 4) No DEWA staff or his or her relatives up to third degree should have ownership or partnership in your company, and your participation in DEWA tenders / RFQs should not constitute a Conflict or perceived Conflict of Interest.
- 5) The offered product and/ or services in the Quotation, shall be conforming and in accordance with DEWA Energy Management Policy & EnMS Manual.

SUPPLIER'S REMARKS :

SUPPLIER'S SIGNATURE AND STAMP